



Hollister Incorporated
2000 Hollister Drive
Libertyville, IL 60048

July 10, 2026

URGENT PRODUCT FIELD SAFETY NOTICE

Onli™ and Infyna™ Hydrophilic Urethral Intermittent Catheter

Dear Customer,

In order to best serve the needs of our customers, Hollister Incorporated, following agreement with the Health Products Regulatory Authority in Ireland (HPRA), is conducting a product alert on the following products:

SKU/REF	Description
82081	Ch08-17 Onli Hydrophilic Urethral Intermittent Catheter
82101	Ch10-17 Onli Hydrophilic Urethral Intermittent Catheter
82121	Ch12-17 Onli Hydrophilic Urethral Intermittent Catheter
82141	Ch14-17 Onli Hydrophilic Urethral Intermittent Catheter
82084	Ch08 40cm Onli Hydrophilic Urethral Intermittent Catheter
82104	Ch10 40cm Onli Hydrophilic Urethral Intermittent Catheter
82124	Ch12 40cm Onli Hydrophilic Urethral Intermittent Catheter
82144	Ch14-40cm Onli Hydrophilic Urethral Intermittent Catheter
82164	Ch16 40cm Onli Hydrophilic Urethral Intermittent Catheter
85124	CH12 40cm Onli Coude Hydrophilic Urethral Intermittent Catheter
85144	CH14 40cm Onli Coude Hydrophilic Urethral Intermittent Catheter
85164	CH16 40cm Onli Coude Hydrophilic Urethral Intermittent Catheter

The Competent (Regulatory) Authority of your country has been informed about this communication to customers

Device Use

This product is a flexible tubular device that is inserted through the urethra by male, female and pediatric patients who need to drain urine from the bladder.

Reason for Product Field Safety Notice

The reason for this Product Field Safety Notice is Hollister has identified that select lots of the catheters identified in the above table may exhibit a protrusion caused by a manufacturing defect which might lead to genitourinary trauma or bleeding. This defect is approximately 2 centimeters from the funnel (refer to image #1 and #2 below). Defect highlighted in red circle.

Hollister is conducting a Product Field Safety Notice to make users aware that this protrusion may be present and as a precautionary measure to inspect catheters for this defect prior to use.

Image 1



Image 2



Actions to be Taken by users

Step #1: Verify product lot number against list attached to this Product Field Safety Notice per image below.

Example of the product reference number (REF) and lot number, found on the individual package:



Example of product reference number (REF), found on the control label on the end of the box:



Step #2: If you are in possession of an affected lot, you may continue to use the device but **should inspect each catheter at the time of use for the potential presence of a protrusion located approximately 2 cm from the funnel** (as described in Images 1 and 2). **If a protrusion is identified, the catheter should not be used.** Users should also review and follow the Instructions for Use (IFU) and ensure that the device is opened, handled, and inserted in accordance with the IFU.

In case of serious injury (incident) in relation to your use of the product, please contact healthcare provider and / or your local distributor.

Please follow the steps below:

1. Please complete and return the response form attached as soon as possible, even if you do not have product in your possession and send the form to ComplaintCommunication@Hollister.com.
2. If you have supplied or transferred any potentially affected product to another facility or organization, provide that facility with a copy of this letter immediately.

For further information or to request replacement devices please contact us:

By phone –Please contact one of the following:

- If you are a consumer or healthcare professional, please call: 1.888.808.7456
- If you are a dealer, distributor, Institution, or pharmacy, please call:
1.800.323.4060 option 1

Thank you again for being our customer and for your understanding regarding this matter.

Sincerely,

Christine Ponka

Christine Ponka
Manager, Post Market Surveillance

URGENT PRODUCT FIELD SAFETY NOTICE

Onli™ and Infyna™ 17cm Hydrophilic Urethral Intermittent Catheter Response is Required

Please return completed response form, even if you do not have affected product in your possession, to ComplaintCommunication@Hollister.com.

Affected Products

SKU/REF	Description
82081	Ch08-17 Onli Hydrophilic Urethral Intermittent Catheter
82101	Ch10-17 Onli Hydrophilic Urethral Intermittent Catheter
82121	Ch12-17 Onli Hydrophilic Urethral Intermittent Catheter
82141	Ch14-17 Onli Hydrophilic Urethral Intermittent Catheter
82084	Ch08 40cm Onli Hydrophilic Urethral Intermittent Catheter
82104	Ch10 40cm Onli Hydrophilic Urethral Intermittent Catheter
82124	Ch12 40cm Onli Hydrophilic Urethral Intermittent Catheter
82144	Ch14-40cm Onli Hydrophilic Urethral Intermittent Catheter
82164	Ch16 40cm Onli Hydrophilic Urethral Intermittent Catheter
85124	CH12 40cm Onli Coude Hydrophilic Urethral Intermittent Catheter
85144	CH14 40cm Onli Coude Hydrophilic Urethral Intermittent Catheter
85164	CH16 40cm Onli Coude Hydrophilic Urethral Intermittent Catheter

URGENT PRODUCT FIELD SAFETY NOTICE

Onli™ and Infyna™ 17cm Hydrophilic Urethral Intermittent Catheter Response is Required

YES ☐	I have read and understood the instructions provided in the Product Field Safety Notice.	
YES ☐	N/A ☐	I have inspected affected product and followed instructions provided in the Product Field Safety Notice.
YES ☐	N/A ☐	I have no affected product by this Field Safety Notice in my possession.
YES ☐	N/A ☐	<i>Distributors:</i> I have identified and notified my customers that were shipped or may have been shipped this product, if applicable. For <u>Credit Requests</u> please contact Hollister Customer Service at 1-800-323-4060 (prompt #1) and provide your purchase order number (PO number) at the time of your request.
*If you do not purchase product directly from Hollister, please provide your supplier's name:		
Please Complete Contact Information for Person Completing Response Form:		
Name and title:		Date:
Telephone Number and/or email:		
Facility/Business name (please use full facility name. Do not use abbreviations):		
Address including city / town, state / county, post code, Country		

URGENT PRODUCT FIELD SAFETY NOTICE

Onli™ and Infyna™ 17cm Hydrophilic Urethral Intermittent Catheter

Affected Products and Lots

Product	Lot
82081	5F273
	5F283
	5H293
	5J153
	6C313
82101	5F283
	5G013
	5K253
	6A303
82121	5H283
	5H293
	5K183
	5K193
	5K213
	6A253
	6A263
	6B053
82141	5E063
	5E073
	5E083
	5F193
	5F203
	5F213
	5F223
	6B133

82084	5E293 5G213 5G223 6A223 6A213
82104	5E293 5E303 5G243 5G253 5J063 5J073 5L153 6D203 6E293
82124	5E303 5F033 5F063 5F043 5F053 5J223 5J283 5J243 5K143 6A163 6A153 6B173 6B183 5F073

82144	6A163
	5F113
	5F133
	5I073
	5G083
	5E243
	5E253
	5E233
	5F103
	5I123
	5F143
	5E263
	5E273
	5I113
	5E283
	5F123
	5F153
	5G093
	5G183
	5G173
	5I153
	5I133
	5I143
	5I063
	6A173
	6B243
	5I103
	6B223
	6B233
	6B253
	6B263
	6B213
	6C033
	6C043
	6B273
	6C013
	6C023
	6C053
	6B283

82164	6A223
	6B183
	5F153
	5F163
	5F183
	5F173
	5G103
	5G143
	5G113
	5G153
	6A233
	6C253
	6C243
	6C223
	6C233
	6D303
	6D293
	6D283
85124	5K143
	5K133
	6B193

85144	6B263
	6B273
	6C253
	5K133
	5K123
	5K113
	5K053
	5K043
	5K063
	5K073
	6A113
	6A133
	5K103
	6A123
	6A093
	6A103
	6C283
	6C263
	6C273
	6E013
85164	6D303
	5K133
	6B193
	6A093
	6C283
	5L053
	5L043
	6A083
	5L193
	6A023
	5L223
	6A053
	6C293